



# EC Declaration of Conformity

We herewith declare, in exclusive responsibility, that the instrument

## **Leica CV5030 – Robotic Coverslipper**

was developed, designed and manufactured to conform with the

- Directive 98/79/EC of the European Parliament and of the Council (in-vitro diagnostic medical devices)


The following harmonized standards were applied:

- **EN 14971: 2007**  
Medical devices - Application of risk management to medical devices
- **EN 591: 2001**  
Instruction for use for in vitro diagnostic instruments for professional use
- **EN 980: 2008**  
Symbols for use in the labelling of medical devices
- **EN ISO 13485: 2003 + AC: 2007**  
Medical devices - Quality management systems - Requirements for regulatory purposes
- **EN 61010-1: 2001 + Corr.: 2002 + Corr.: 2004**  
Safety requirements for electrical equipment for measurement, control and laboratory use - Part 1: General requirements
- **DIN EN 61010-2-101: 2002**  
Safety requirement for electrical equipment for measurement, control and laboratory use Part 2-101: Particular requirements for in vitro diagnostic (IVD)
- **EN 61326: 2006 + Corr.: 2008**  
Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements

In addition, the following in-house standards were applied:

- **DIN EN ISO 9001: 2000**  
Quality management systems - Requirements

Leica Biosystems Nussloch GmbH  
Heidelberger Str. 17-19  
69226 Nussloch, Germany  
October 30, 2009

  
Arnd Kaldowski  
President Biosystems Division