

**EC Declaration of Conformity**  
**EG Konformitätserklärung**



Leica Microsystems CMS GmbH  
Ernst-Leitz-Straße 17-37  
D-35578 Wetzlar

We hereby declare that the device described below, both in its basic design and construction and in the version marked by us, conforms to the relevant safety- and health- related requirements of the appropriate EU directives.

This declaration shall cease to be valid if modifications are made to the device without our approval.

Hiermit erklären wir, daß nachfolgend bezeichnetes Gerät aufgrund seiner Konzipierung und Bauart sowie in der von uns in Verkehr gebrachten Ausführung, den einschlägigen grundlegenden Sicherheits- und Gesundheitsanforderungen der EU Richtlinien entspricht. Bei einer nicht mit uns abgestimmten Änderung des Gerätes, verliert diese Erklärung ihre Gültigkeit.

**Product / Bezeichnung:**

**Microscope / Mikroskop**

Model:/ Gerätetyp:

**DMI4000 B, CTR4000**

Identification No / Gerätenummer :

**11 888 315, 11 888 316, 11 888 317, 11 888 318,  
11 888 290, 11 888 291, 11 888 164**

EU directives / EU Richtlinien:

In vitro Diagnostic Directive : 98/79/EG  
In vitro Diagnostica : 98/79/EG

Harmonised standards applied /  
Angewandte harmonisierte Normen :

EN 61010-1 : 2001  
EN 61010-2-101 : 2002  
EN 61000-3-2 : 2001  
EN 61000-3-3 : 2001  
EN 61326:1997 + A1 :1998 + A2:2001  
EN 14971 : 2001

ppa. Dr.L.Remer,  
Leiter Entwicklung u. Technologieplanung Mikroskopie

Wetzlar, den 20.05.2005