



Certificate of Compliance

Certificate: 2084242

Master Contract: 217333

Project: 2084242

Date Issued: 2008/10/22

Issued to: Leica Biosystems Nussloch GmbH

Heidelbergerstrasse 17-19

Nussloch, 69226

Germany

Attention: ROBERT GROPP

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Eknath Desai; C.E.T.

Authorized by: George Tranquada, C.E.T.
Product Group Manager

PRODUCTS

CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

CLASS 8721 05 - LABORATORY EQUIPMENT - Electrical

PRINTER - Models, Leica IPC, Leica IPS, Sakura Tissue-Tek Auto Write Cassette Printer, Sakura Tissue-Tek Auto Write Slide Printer, cord connected, rated , 100 - 120V/ 200 - 240V, 50 - 60Hz, $\frac{4}{2,8}$ A

CSA C22.2 No. 1010.1 (Ver 2.0) CERTIFICATION REPORT SUPPLEMENT (DESIGN MANUAL ISSUED UNDER PREVIOUS APPLICATION) IS AN INTEGRAL PART OF THIS REPORT.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 2084242

Master Contract: 217333

Project: 2084242

Date Issued: 2008/10/22

APPLICABLE REQUIREMENTS

CAN / CSA - C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use - Part 1: General Requirements

CONDITIONS OF ACCEPTABILITY

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.