



America

CERTIFICATE

No. U8 15 07 56257 128

Holder of Certificate: Leica Microsystems CMS GmbH

Ernst-Leitz-Straße 17-37
35578 Wetzlar
GERMANY

Production Facility(ies):

56257

Certification Mark:



C US

Product:

Laboratory equipment
(Microscope)

Model(s):

LMD 6, LMD 7
LMD 6: 11888871 (IVD), 11888869 (IVD),
11888868, 11888870
LMD 7: 11888876 (IVD), 11888877 (IVD),
11888875, 11888878

Parameters:

Rated voltage:	12 VDC
Protection class:	III
Remark:	Supplied via control unit CTR6 LED or control unit CTR6

Tested according to:

UL 61010-1:2012-05
UL 61010-1:2004/R:2008-10
IEC 61010-2-101(ed.1)
CAN/CSA-C22.2 No. 61010-1:2012-05
CAN/CSA-C22.2 No. 61010-1:2004-07
CAN/CSA-C22.2 No. 61010-2-101:2004-07

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC Guide 67. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713059304-000

Date, 2015-07-30

Page 1 of 1

